Waters

MODIFYING WATERS AMDS FOR IS™ METHOD DEVELOPMENT

Michael D. Jones and Kate Evans, Ph.D Waters Corporation

The Waters® Automated Method Development System (AMDS) provides an alternative to traditional, inefficient, manual and often unreliable trial-and-error processes typically used to develop and optimize HPLC methods. AMDS combines the precision, accuracy and reliability of the Alliance® HPLC System with the ease and flexibility of Empower™ Software configured with the automated method development decision manager. The System provides an automated and efficient way of developing, optimizing and transferring high quality, robust HPLC methods. The possibility of combining AMDS with the latest Waters Intelligent Speed (IS[™]) Column technology, provides even greater opportunities to further optimize chromatographic separation speed, thus increasing productivity.

AMDS is currently configured with standard analytical column dimension algorithms embedded in the internal logic that calculates chromatographic starting conditions. With the increasing popularity of *IS* technology, more scientists want to incorporate *IS* Columns into their methods to leverage the separation speed for productivity gains. Proper modifications must be made within the AMDS software, and the Alliance 2695¹ and 2996 PDA² must be optimized to achieve the lowest system volume possible. This technical note will explain the procedures and configuration changes needed to use Waters AMDS to develop methods with *IS* Columns, eliminating the need to first develop methods on analytical column dimensions before converting to *IS* Columns.

Experimental

In order implement *IS* Column technology, we must first understand how AMDS calculates flow rates and gradient run times. AMDS version 1.2 determines flow rate as follows:

 $F = (1/L) * (Dc^2) * (Dp^2) * (Pmax/1000) * Pf *Sf$

Where:

F = flow rate in mL/min
L = column length in cm
Dc = column id in cm
Dp = packing particle size in microns
Pmax = maximum pressure in psi
Pf = pressure factor
Sf = solvent factor: (Water = 2.0, ACN = 3.75, MeOH = 2.0)

The pressure factor is predetermined experimentally with conventional analytical column dimensions. Currently, this value is 0.5 for both Waters XTerra™ and Symmetry® Columns. This pressure factor considers *column* backpressure, but not *system* backpressure. Observed HPLC system backpressure is negligible at optimum flow rates for conventional analytical columns. The optimal flow rate for IS columns is much higher (approximately 3.0 - 4.0 mL), and therefore system backpressure is a factor. Currently, AMDS does not take this pressure into account when proposing starting conditions. The potential increase of 1000 psi contributed by system backpressure must be compensated for in the AMDS proposal screen. The flexibility of AMDS allows the user to make these flow rate modifications in the "Propose Data Run" screen as depicted in Table 1. Note the "User" modifications are documented with a dot (•) to signify a change to the AMDS proposed starting conditions.

Step: Propose Data Runs			
Condition	Step	Proposed Value	User
Low temperature	Alİ	35 °C	
High temperature	All	50 °C	
Injection Volume	All	20 µl	
Maximum run time	All	5 min	
Minimum resolution	All	1.5	
Maximum pressure	All	3000 psi	
A Solvent	All	10 mM ammonium bicarbonate, pH 9	7
B Solvent	All	acetonitrile	
Flow Rate	ACN	3.0 ml/min	•
Short gradient time	ACN	1.5 min	•
Long gradient time	ACN	3.0 min	•
Low %B	ACN	5 %	
High %B	ACN	85 %	

echnicalNOT

Table 1. Typical "Proposal Screen" modifications when using IS Columns. This is a table taken from the AMDS message log file. Modifications are highlight in red.

The other starting conditions that may have to be modified are the short and long gradient times, where gradient time is depicted in Drylab as tG (time of Gradient). The long gradient is advised to be approximately a factor of 3 greater than the short gradient in order to achieve an optimal Drylab resolution plot. The calculation for tG is as follows:

tG = 0.01 * (k*) * (Max %B - Min %B) * (0.5 *(column length) *

(Column diameter)^2) * s-value/flow rate

Vaters

WATERS CORPORATION 34 Maple St., Milford, MA 01757 U.S.A. T: 508 478 2000 F: 508 872 1990 www.waters.com

Where:

k* for short gradient = 2 (defaults in the AMDS configuration screen)

k* for long gradient = 6 (defaults in the AMDS configuration screen)

S-value = 5 (a value related to molecular weight of the samples. 5 is suitable for small molecules)

tG = gradient time

One can see that flow rate is an important factor in the equation. Since the flow rate (F) has been modified the gradient time (tG) also needs to be modified. As a rule of thumb, it is best to choose short and long gradient that are a factor of 3 apart, and allow for gradient profiles that provide enough time to collect data for Drylab's predictions.

References

 Jenkins, T., et al "Alliance 2695 Separations Module: Optimization and Performance with 4.6mm
 I.D. Intelligent Speed (ISTM) Columns." Waters Technical Note – 720000723EN

 Jenkins, T., et al "2996 Photodiode Array Detector: Optimization for Intelligent Speed (*IS™*) Columns." Waters Technical Note – 720000722EN

Related Information

Diehl, D., et al "Intelligent Speed: Theory and Practice for Increasing Throughput" WA31779

For Complete St Confidence

Waters, Alliance, IS, XTerra, Symmetry and Empower are trademarks of Waters Corporation. All other trademarks are property of their respective owners. ©2004 Waters Corporation Produced in the USA November 2004 720001045EN LW-PDF

