

Extraction of Phospholipids from Plasma Using Ostro Pass-through Sample Preparation

Mark Ritchie, Mainak Mal, and Stephen Wong
Waters Pacific, Singapore

APPLICATION BENEFITS

- Highly reproducible selective enrichment of certain classes or total phospholipid extraction.
- The 96-well format is compatible with automation technologies.
- Requires lower volumes of solvents than liquid extraction methods.

WATERS SOLUTIONS

Ostro™ Pass-through
Sample Preparation Products

Xevo® TQ-S Mass Spectrometer

ACQUITY UPLC® System

ACQUITY® BEH HILIC Columns

KEY WORDS

Plasma, phospholipid extraction, Ostro

INTRODUCTION

The common goals of many “omics” are to identify disease molecular markers and to understand mechanisms of disease which can then suggest possible drug targets. Typically, circulatory body fluids are studied as they have significant contact with affected organs and tissues, and so are sensitive to the body’s specific responses to a pathological condition.

In lipidomic analysis, blood plasma phospholipids are usually studied following solvent extraction.^{1,2} Solid-phase extraction techniques have also been used in the past. Aminopropyl silica columns were used to analyse lipids, but they either focussed more on fatty acids,³ or were class selective depending on the elution conditions used.⁴

The Ostro 96-well Sample Preparation Plate is a pass-through device designed to capture and remove the highly abundant phospholipids as part of sample preparation for the analysis of small molecules during routine bioanalysis. In the generic protocol, the lipids are retained by the Ostro Pass-through Plate, and no attempt is made to release them. Modifying the loading and elution conditions allows this same affinity to be exploited for both selective enrichment of certain classes or for total phospholipid extraction.

EXPERIMENTAL**ACQUITY UPLC conditions**

Column:	ACQUITY BEH HILIC 1.7 µm, 2.1 x 100 mm
Column temp.:	30 °C
Mobile phase A:	Acetonitrile/water (95:5) with 10 mM ammonium acetate, pH 8.0
Mobile phase B:	Acetonitrile/water (50:50) with 10 mM ammonium acetate, pH 8.0
Gradient:	0–20% B for 10 min
Flow rate:	500 µL/min
Injection volume:	3.0 µL, partial loop

Xevo TQ-S MS conditions

Ionization mode:	ESI, +/- switching
Capillary voltage:	3.0 kV
Desolvation temp.:	450 °C
Desolvation gas:	1000 L/hr
Source temp.:	150 °C
Collision cell pressure:	3.6 x 10 ⁻³ mBar

Sample preparation

Human plasma was obtained either through volunteers at the Centre for Life Sciences, National University of Singapore, or from a commercial sample (Sigma, p/n P9523).

Extraction Protocol

1. Place the Ostro Plate onto a 2 mL collection plate, and position in the vacuum manifold.
2. Load 100 µL of plasma directly into the wells of the Ostro Pass-through 96-well Plate using a micropipette.
3. Add 800 µL of ethanol to the wells and thoroughly mix by aspirating the samples 10 x using the micropipette.
4. After mixing, apply 15" Hg of vacuum until the solvent drains completely (approx. 1 minute). A higher vacuum setting can be used if flow does not occur at 15" Hg.
5. Repeat steps 3 and 4, and label the collected volume as the "flow through" fraction.
6. Remove the collection plate containing the flow through fractions from the vacuum device and replace with a fresh plate to collect the eluate fraction.
7. Add 800 µL of elution solvent (4.5:4.5:1 Chloroform:Methanol:Triethylamine) to the wells.
8. Apply a vacuum of 15" Hg until the solvent drains completely (approx. 1 minute).
9. Repeat steps 6 and 7, and label this collected volume as the "eluate" fraction.
10. Dry both flow through and eluate fractions (around 1.5 hours using a nitrogen evaporator).
11. Reconstitute the eluate (E) and flow through (FT) fractions in 200 µL of 1:1 (v/v) Chloroform:Methanol.

Note: Take care to ensure all the pipette tips are properly positioned in the sample wells to prevent any loss or mixing of samples.

For comparison, aliquots of the same human plasma were extracted using the Bligh and Dyer method.¹

Due to the large dynamic range in intensities observed between the phospholipid classes, the samples were analysed at two concentrations; undiluted and diluted 1/100 with acetonitrile. A custom MRM method was used with positive and negative switching and MRMs timed to coincide with the elution times of specific lipid classes¹. A summary of this method is shown in Table 1 below.

Lipid Class	MS Polarity	Number of Transitions	Acquisition time of MRM (mins)
LPCs	+	11	6–10
Cer	+	34	0–5
SMs	+	19	5–8.5
PCs	+	48	4–7
LPEs	-	11	4–10
LPIs	-	11	4–10
PIs	-	27	1.5–6
PEs	-	34	1.5–6
PGs	-	20	0–3

Table 1. Summary of MRM transitions by class used in the analysis.

Intensities for each species were automatically extracted as area counts using a pre-programmed TargetLynx™ method. A mean intensity threshold of 1000 area counts was applied to the data.

As an experimental control when comparing the two sample preparation methods, exogenous lipids were spiked into the samples immediately prior to analysis. These were used for normalisation prior to calculating the relative extraction efficiency of a method or fraction. The spiked lipids and their concentrations are shown below in Table 2.

Lipid standard	Final concentration in sample (ng/μL)
DMPA	0.5
DMPG	1.0
DMPE	2.5
DMPS	1.0
diC8PI	1.0
DMPC	5.0
SM	1.0
C ₁₇ Cer	1.0
C ₈ GluCer	0.5

Table 2. Exogenous lipid spikes used for normalisation during the method comparison.

RESULTS AND DISCUSSION

Comparing extraction efficiencies using Ostro Pass-through Sample Preparation to the Bligh and Dyer method

Extractions were carried out on mixed volunteer human plasma in duplicate, with analytical duplicates. A threshold of 1000 counts from the appropriate MRM must be exceeded for the detection of a lipid. Peak areas for each lipid species were normalized as a ratio to a post extraction internal spike of an exogenous lipid of the same class. These ratios were then averaged across replicates in the same fraction/method to give a single normalized value for each lipid in that fraction. The values for lipids from the Ostro Pass-through and eluate fractions were divided by the values for the same lipids from the Bligh and Dyer method to give an extraction efficiency expressed as a ratio for each lipid. The average extraction efficiencies of each class are shown in Table 3.

Lipid Class	Number of lipids detected			Extraction Efficiency of Ostro vs BD	
	BD	Ostro E	Ostro FT	Ostro E	Ostro FT
SM	17	17	0	1.00	0.00
LPC	9	9	0	1.00	0.00
PC	47	47	0	1.01	0.00
Cer	13	3	13	0.03	0.95
PG	15	5	15	0.06	1.37
LPE	10	6	10	0.07	0.60
PE	28	23	28	0.14	0.99
LPI	0	4	5	n/c	n/c
PI	22	18	22	0.16	1.14

Table 3. Comparison of efficiency of sample preparation methods on volunteer human plasma. The number of lipids observed (exceeding 1000 area counts) in each class are shown for both Ostro and Bligh and Dyer methods. The extraction efficiency is expressed as a ratio of normalised intensity of a given species extracted by Ostro Pass-through sample preparation relative to that of the Bligh and Dyer method. Green shading highlights the fraction containing the majority of the area counts for that lipid class. Key: BD: Bligh and Dyer method,¹ Ostro E: Eluate fraction from Ostro Pass-through sample preparation, Ostro FT: Flow Through fraction from Ostro Pass-through sample preparation, n/c: no comparison possible as species were not observed by one method.

Sphingomyelins (SMs), phosphatidylcholines (PCs), and lyso-phosphatidylcholines (LPCs) are completely retained on the Ostro plate and only appear in the eluate. The same phospholipid species were observed by both extraction methods at the same intensities.

Ceramides (Cer) and phosphatidylglycerols (PGs) appear to have little affinity for the solid phase under these conditions and are found in the flow through fraction, with very small amounts retained and recovered in the eluate fraction. While the same PG lipids were identified, extraction using the Ostro Pass-through 96-well Plate appears to increase the area counts of the PGs compared the Bligh and Dyer method. Phosphatidylethanolamines (PEs) and phosphatidylinositols (PIs) follow a similar behaviour but slightly more lipid is retained by the Ostro plate and appears in the eluate, however, the extraction efficiency appears to be the same as the Bligh and Dyer method. The same lyso-PE (LPE) species are observed by both extraction methods, but it appears that the recovery from the Ostro Pass-through Plate is about 60% of the Bligh and Dyer method.

Extraction of lyso-phosphatidylinositols (LPIs) on the other hand is greatly enhanced by Ostro sample preparation, with 5 additional lipid species being observed using the Ostro Pass-through Plate than by the Bligh and Dyer method. The majority of the LPI (90%) does not bind to the Ostro Pass-through Plate.

In general, SMs, PCs and LPCs bind to the Ostro Pass-through Plate and are only eluted with 4.5:4.5:1 Chloroform:Methanol: Triethylamine. All other classes have minimal interaction with the plate under the conditions used and pass through in the flow through fraction.

Reproducibility of the Ostro Pass-through Phospholipid Extraction Method

The Ostro Pass-through Plate was used to extract 14 aliquots of commercially available human plasma. The flow through (FT) and eluate (E) fractions were analysed separately in triplicate for each extraction. Percentage relative standard deviations (%RSDs) of the area counts were calculated for each lipid species across all replicates in the data set and listed in Table 4.

Lipid Class	MS polarity	Fraction	Number lipids detected	%RSD of area counts
Cer	+	FT	13	13.3
LPC	+	E	9	14.1
SM	+	E	17	11.3
PC	+	E	46	9.1
LPE	-	FT	10	10.9
LPI	-	FT	7	22.2
PI	-	FT	25	12.5
PE	-	FT	29	16.2
PG	-	FT	18	10.3
All measurements				16.5

Table 4. Percentage Relative Standard Deviations (%RSD) for intensity measured in area counts from 14 replicate extracts of commercially available plasma using Ostro Pass-through sample preparation. Key; FT: Flow Through fraction, E: Eluate fraction.

The %RSD shown are influenced by ion intensity, with the less abundant classes contributing more to the overall error in this experiment. The error of the analytical (LC-MS) method has been shown to be 5% at the same intensity threshold.⁵

CONCLUSIONS

The Ostro Pass-through sample preparation method for phospholipids is a simple and reproducible method.

The number and amount of the phospholipid species observed in plasma by this method compares very well to those observed by the standard extraction method (Bligh and Dyer), with the Ostro Pass-through Plate having significant benefits in recovering LPIs and PGs.

This technique is suitable for both small scale and large scale lipidomic studies. As the Ostro Pass-through Plate uses a 96-well format and requires no centrifugation, it can potentially be automated with a laboratory robotics system, which cannot be done with the liquid extraction method. The Ostro Pass-through Sample Preparation Method also uses significantly less solvent than liquid extraction methods.

This technique can be used either as a general all-class extraction method for phospholipids, or it can be used selectively. For example, analyzing only the flow-through fraction allows low abundance species to be studied without dynamic range and ion suppression issues caused by the highly abundant PCs.

References

1. Bligh, E.G and Dyer, W.J.; *Canad. J. Biochem. Physiol* (1959) 27, 911-917.
2. Zhao, Z and Xu, Y; *J. Lipid Res* (2010) 51, 652-659.
3. Burdge *et al.*; *British Journal of Nutrition* (2000), 84, 781-787.
4. Kim *et al.*; *J. Lipid Res* (1990), 31, 2285-2289.
5. UPLC BEH HILIC: "The Preferred Separation Chemistry for Targeted Analysis of Phospholipids", Mark Ritchie, Mainak Mal, and Stephen Wong. Waters Application Note no. 720004219EN, January 2012.

Waters

THE SCIENCE OF WHAT'S POSSIBLE.®

Waters, The Science of What's Possible, Xevo, ACQUITY, and ACQUITY UPLC are registered trademarks of Waters Corporation. TargetLynx and Ostro are trademarks of Waters Corporation. All other trademarks are the property of their respective owners.

©2014 Waters Corporation. Produced in the U.S.A. November 2014 720004201EN AG-PDF

Waters Corporation
34 Maple Street
Milford, MA 01757 U.S.A.
T: 1 508 478 2000
F: 1 508 872 1990
www.waters.com