


CLICK ON THE UNDERLINED BLUE TEXT FOR DETAILS ON THE PRODUCTS USED IN THIS APPLICATION

## TEST CONDITIONS

Column: ACQUITY UPLC® BEH Amide,  
2.1 x 150 mm, 1.7 µm  
Part Number: [186004802](#)  
Isocratic Mobile Phase: 95/2.5/2.5 MeCN/IPA/H<sub>2</sub>O with 10 mM  
CH<sub>3</sub>COONH<sub>4</sub> and 0.01% NH<sub>4</sub>OH, pH 9.0  
Flow Rate: 0.2 mL/min  
Injection Volume: 5.0 µL (PLNO)  
Sample Concentration: 10 µg/mL  
Sample Diluent: 75/25 MeCN/MeOH with 0.2% HCOOH  
Column Temperature: 25 °C  
Weak Needle Wash: 95/5 MeCN/H<sub>2</sub>O  
Detection: UV @ 245 nm  
Sampling Rate: 20 points/sec  
Filter Time Constant: 0.2  
Instrument: Waters ACQUITY UPLC with  
ACQUITY UPLC PDA Detector

## STRUCTURE



Thiourea

