#### CLICK ON THE UNDERLINED BLUE TEXT FOR DETAILS ON THE PRODUCTS USED IN THIS APPLICATION

## TEST CONDITIONS

### Chromatographic Conditions

Column: ACQUITY UPLC® BEH Amide

 $2.1 \times 100$  mm,  $1.7 \mu m$ 

Part Number: <u>186004801</u>

Mobile Phase A: 80/20 MeCN/H<sub>2</sub>0 with 0.2%

triethylamine [TEA]

Mobile Phase B: 30/70 MeCN/H<sub>2</sub>0 with 0.2%

triethylamine [TEA]

Flow Rate: 0.13 mL/min

Gradient: 10 minute gradient, 75%-45%

MeCN (w/0.2% TEA) with 25 minute

re-equilibration

Time	Profile	
(min)	%A	%B
0.00	90.00	10.00
10.00	30.00	70.00
10.01	90.00	10.00
35.00	90.00	10.00

Injection Volume:  $1.3 \mu L (PLNO)$ 

Sample Concentration: Standards at 1 mg/mL, beer at 100%

(No dilution)

Sample Diluent: 50/50 MeCN/H<sub>2</sub>0

Column Temperature: 35 °C


Strong Needle Wash: 20/80 MeCN/ $H_2$ 0 (800  $\mu$ L) Weak Needle Wash: 75/25 MeCN/ $H_2$ 0 (500  $\mu$ L)

Seal Wash: 50/50 MeCN/H<sub>2</sub>0

Instrument: Waters ACQUITY UPLC with ELSD

## **ELSD Conditions**

Gain: 200
Pressure: 40 psi
Drift Tube Temperature: 40 °C
Nebulizer: Cooling
Data Rate: 10 pps
Filter Time Constant: Normal



# COMPOUNDS

