

Acyclovir



HPLC Method

Column: Symmetry® C₈, 3.9 x 150 mm, 5 µm
Part number: WAT046970
Mobile phase: 0.1% glacial acetic acid (aq), pH 3.3
Flow rate: 1.0 mL/min
Injection volume: 5 µL of 100 µg/mL sample
Detection: UV @ 254 nm

USP Tailing Factors

1. 1.1

