

Cytochrome c Tryptic Digest - Recovery Study



HPLC Method

Column: Symmetry300™ C₁₈, 4.6 x 50 mm, 5 µm
Part number: WAT106209
Mobile phase A: 0.1% TFA in Water
Mobile phase B: 0.1% TFA in Acetonitrile
Gradient:

Time (min)	%A	%B
0	100	0
45	70	30

Flow rate: 1.0 mL/min
Injection volume: 20 µL
Temperature: 35°C
Detection: UV @ 214 nm

