



Diphenoxylate HCL

HPLC Method

Column: Symmetry® C₁₈, 3.9 x 150 mm, 5 µm
Part number: WAT046980
Mobile phase: 50 mM potassium phosphate, pH 6.0/acetonitrile 30:70
Flow rate: 1.0 mL/min
Injection volume: 10 µL of 100 µg/mL Diphenoxylate hydrochloride
Detection: UV @ 210 nm

USP Tailing Factors

1. 1.3



1. Diphenoxylate HCL